



जीव विज्ञान संस्थान

INSTITUTE OF LIFE SCIENCES
(An Autonomous Institute of Department of Biotechnology, Govt. of India)
NALCO SQUARE, BHUBANESWAR-751 023

Corrigendum

No.: IV- 263 -S&P/2017-18/ 250 /SLT/SE/ Cgdm/ILS dated 25.01.2018

Corrigendum to the Sealed Limited Tender No. IV- 263 -S&P/2017-18/ 154 /SLT/SE/ILS dated 17.01.2018 for Supply, Installation & Satisfactory Demonstration of **Benchtop Centrifuge**

The specifications of the equipment is modified as follows:

Sl.	Modified specification
1	<u>Benchtop centrifuge</u> : The high Capacity multipurpose centrifuge of Max Speed: ~20,000g or greater, with fixed-angle rotors for microcentrifuge tubes, 50ml, 15ml centrifuge tubes (with adaptors for conical or round bottom tubes), swing out rotor with adaptors for bottle, 50ml/15ml falcon tubes, blood collection tubes/vacutainers. Buckets should have aerosol-tight caps with biosafety features. The rotors should be lightweight and rust free. The weight of the rotors need to be mentioned and rotors should have biosafety features with autoclavable lids and containers. The inner body should be rust free. Temperature range: - 8°C to 40°C, fast temperature function for rapid cooling and standby cooling function, condensation drainage system, low noise < 56 dB(A), acceleration & breaking ramp 10/10 timer and timer should start after set RPM is reached, short spin key option, values can be changed during centrifugation, automatic rotor recognition & imbalance/error detection, plate buckets for MTP with 3 years of warranty from the date of installation.

Bidders already dispatched their bids may revise or withdraw their bids as per the above corrigendum. All other terms & conditions of the above Sealed Limited Tender remain unchanged.

Sd/-
Stores & Purchase Officer