

**TECHNICAL VERIFICATION REPORT (01 PAGE)****Purchase of Scientific Equipment Vide Sealed Limited Tender Call Notice No.: IV-263-S&P/2017-18/1409/SLT/SE/ILS dated 30.06.2017****Items Sl. No.: 01 (One) : Biosafety Cabinet**

Sl. No.	Name of the Firm	Principal Name	Make/ Model	Qualified/ Disqualified	Reason of Disqualification
1.	M/s Aplha Linear, Bangalore	Own make	Make: Own make, Model: BS-2422-A <sub>2</sub>	Disqualified	Time setting function & continuous HEPA filter monitoring system not complied <ul style="list-style-type: none"> <li>• Static pressure measurement missing</li> <li>• Permanent backup HEPA filter missing</li> </ul>
2.	M/s Tecpharma, Hyderabad	H Biomedical Limited, Hong Kong	Make: Haier Model: HR40II A <sub>2</sub>	Qualified	
3.	M/s Bio-technika, Bhubaneswar	M/s Microfilt India Pvt. Ltd., Pune	Model: MFI BIO 4X <sub>2</sub>	Disqualified	<ul style="list-style-type: none"> <li>• Double sided DC motor blower not complying</li> <li>• Permanent backup HEPA filter missing</li> <li>• Air circulation measurement is in approximation</li> </ul>
4.	M/s Suravi Instrumentation Pvt. Ltd., Kolkata	M/s Thermo Electron LED GmbH, Germany	MSC Advantage 1.2 of Thermo Heraeus	Qualified	Provided company agrees to provide original stand along with in the same price
5.	M/s Clean Air System, Chennai	Own make	Cleanair® Biosafety cabinet Class II, Type A <sub>2</sub> Model: CBS1200 A <sub>2</sub>	Disqualified	<ul style="list-style-type: none"> <li>• Main filter ULPA is not there</li> <li>• Blower is not double sided with DC motor</li> <li>• Florescence</li> </ul>
6.	M/s Labchem & Labortenik Instruments, Kolkata	M/s Esco Micro Pte. Ltd., Singapore	Model: LA <sub>2</sub> -4A1-E	Disqualified	<ul style="list-style-type: none"> <li>• Exhaust filter indented is HEPA</li> <li>• Dimensions are much bigger than indented (that much space not available in cell culture)</li> </ul>

**Items Sl. No.: 02 (Two) : Contamination Monitor**

Sl. No.	Name of the Firm	Principal Name	Make/ Model	Qualified/ Disqualified	Reason of Disqualification
1.	M/s Nucleonix Systems Pvt. Ltd., Hyderabad	Own make	Digital contamination monitor Type: CM710N	Disqualified	<ul style="list-style-type: none"> <li>• No α-monitoring facility</li> </ul>
2.	M/s Berthold Technologies (India) Pvt. Ltd., Chennai	M/s Bethhold Technologies GmbH & Co. KG, Germany	LB124 SCINT Alpha-Beta contamination monitor	Qualified	