

**TECHNICAL VERIFICATION REPORT (01 PAGE)**

**Purchase of "Inverted Microscope" Vide Sealed Tender Call Notice No: IV-263-S&P/2017-18/153/SLT/SE/ILS dated 17.01.2018**

Name of the Firm	Principal Name	Make/Model	Qualified/ Disqualified	Reason of Disqualification
M/s DSS Imagetech Pvt. Ltd., Kolkata - 700 029.	M/s Olympus Corporation, Japan	Model: CKX53 with EM 10 Camera	Qualified	
M/s Systems & Controls, New Delhi - 110 058.	M/s Motic Hong Kong Limited Hong Kong	AE31E with Moticam 1080	Disqualified	1. No details or LED illumination & hour in the brochure 2. No details of camera & line cell imaging in the brochure
M/s Radical Scientific Equipments Pvt. Ltd., Ambala Cantt. - 04	Own make	Radical Model RTC-7 with ProCam True Chrome 5mp HD Camera	Disqualified	1. No details as LED working hours in the brochure 2. Camera details such as megapixels not mentioned
M/s Micro Instrument Company, Ambala Cantt. - 01, Haryana.	Make: Miko NSIC Firm	Model: Mi - IBM. L43 + Mi. HD5000	Disqualified	1. No details as LED hours mentioned in the brochure 2. Objective phase rings information not given • Camera 3 mp
M/s Invitrogen BioServices India Pvt. Ltd., Kolkata - 91.	M/s Life Technologies Holdings Pte Ltd., Singapore	EVOSXL CORE Base (AMEXI000) Microscope System with color camera (Sony Series CMOS)	Disqualified	
M/s ASE Instruments Agency Pvt. Ltd., Kolkata - 700 010.	M/s Carl Zeiss Microscopy GmbH, Germany	Model: PrimoVert with AxioCam 105 Color Alternatively AXIO Cam Erc5S	Disqualified	• Camera Axio 105 color does not have SD card slot & the alternative camera ERC5S does not have required fps for line cell imaging
M/s Leica Microsystems, Division of DHR Holding, Mumbai.	Leica Mikrosysteme Vertrieb GmbH, Germany	DMi1 Leica with MC 170	Qualified	Conditionally qualified. Decision is pending about acceptance of their bid

**APPROVED**

*(Signature)*

**DIRECTOR**

(An Autonomous Institute of  
Dept. of Biotechnology, Govt. of  
INSTITUTE OF LIFE SCIENCES  
Bhubaneswar-751 023, Odisha)

*(Handwritten signature)*