

TECHNICAL VERIFICATION REPORT (01 PAGE)

Purchase of "Ultrasonic Processor / Ultrasonic Probe Sonicator " Vide Sealed Limited Tender Call

Notice No: IV-263-S&P/2017-18/747/SLT/SE/ILS dated 23.03.2018

Sl. No.	Name of the Firm	Principal Name	Make/ Model	Qualified/ Disqualified	Reason of Disqualification
1.	M/s Genetix Biotech (Asia) Pvt. Ltd., New Delhi.	Own make	GX-Sonic-950WT	Disqualified	<ul style="list-style-type: none">• High power output (950W)• Sample volume not more than 600ml (need upto 1L)• EMD part is not clear
2.	M/s Unigenetics Instruments Pvt. Ltd., Delhi - 91.	M/s Ningbo Yinzhou Sjalab Equipment Co. Ltd.	SJIA LAB SKL-750	Disqualified	<ul style="list-style-type: none">• No provision for automatic amplitude compensation (not mentioned in brochure)• No authorization from the manufacturer
3.	M/s Interlab Instruments & Equipments, Kolkata	M/s Thermo Fisher Scientific Inc., USA	FB-705-220	Qualified	
4.	M/s Tecpharma Systems Inc., Hyderabad - 72.	M/s H Biomedical Ltd., Hong Kong	HPS-750 IINA	Disqualified	<ul style="list-style-type: none">• Frequency is higher (25 KHz) Need 20 KHz
5.	M/s Inkarp Instruments Pvt. Ltd., Hyderabad - 07.	M/s Sonic & Materials Inc., USA	SONICS - USA VCX-750	Qualified	
6.	M/s Athena Technology Thane	Own make	ATP-750	Disqualified	<ul style="list-style-type: none">• Wrongly sample volume max is only 900 ml (need upto 1L)• EMD not submitted
7.	M/s Oscar Ultrasonics Pvt. Ltd., Mumbai - 72.	Own make	Sona Pros PR-1000mp	Disqualified	<ul style="list-style-type: none">• No provision for automatic amplitude compensation
8.	M/s Albha Instruments Chennai - 24.	M/s Q Sonica LLC., USA	Qualified		
9.	M/s Chromi Life Inc., Cuttack - 08.	M/s Ningbo Yinzhou Sjalab Equipment Co. Ltd., China	SJIA - 750W Ultrasonic Frequency (KHz)	Disqualified	<ul style="list-style-type: none">• No provision for automatic amplitude compensation (not mentioned in brochure)